

EC Declaration of Conformity

Manufacturer :

Televic Healthcare N.V.

Leo Bekaertlaan 1

B-8870 Izegem

Belgium

Phone : +32 51 30 30 45

declares that following products:

Smart Units:

50.20.0020 Smart presence unit green red

50.20.0040 Smart presence unit green, yellow, white

50.10.0010 Smart call unit red button

50.10.0020 Smart call unit pull cord

50.12.0030 Smart single outlet call/cancel

50.12.0040 Smart double outlet call/cancel

50.14.0020 Smart call unit blue/yellow

50.30.0010 SIP addressable digital intercom unit

Smart Units with connections for passive units:

50.13.0010 AQURA smart ward controller module

50.13.0031 AQURA overdoor light with int. room hub

50.13.0020 AQURA smart room hub

Passive units (auxiliary devices):

50.10.0210 Passive call unit red button

50.10.0220 Passive call unit pull cord

50.12.0210 Passive single outlet with call button with 2 isolated drivers

50.12.0260 Passive single outlet with call button with 4 isolated drivers

50.12.0200 AQURA passive single outlet without button

50.20.0220 Presence unit with call button

50.20.0210 Passive unit with green button and buzzer

50.05.0020 Passive overdoor light with 4 fields and 5 colours

are in conformity with the following requirements :

Directives:

- 2014/30/EU - EMC Directive
- 2011/65/EU - ROHS

Norms:

- EN 55032:2015+A11:2020 (emissive)
- EN 55035:2017+A11:2020 (immunity)
- EN 50581 : 2012

National standards or other specifications (for smart units):


- DIN VDE 0834-1 Specifications for call systems in hospitals, nursing homes and similar institutions.

Quality assurance

- The quality management system of Televic is ISO 9001:2015 certified for its development, production, service and installation of this product.

Izegem, March 2022

T E L E V I C HEALTHCARE N.V.

A handwritten signature in black ink, appearing to read 'J. Gesquiere', with a long horizontal stroke extending to the left.

John Gesquiere
R&D Director